

Certified Reference Values



Production date: 29-07-2024

Batch: 31 - 2024

Type	Unhomogenised past. milk	Unhomogenised past. cream										
	Buttermilk	6,5%	8%	9,5%	11%	12%	14%	18%	32%	38%	46%	50%
Fat (g/100 g) ≤8%+9,5%: ISO 1211/IDF 1:2010 ≥9%: ISO 2450/IDF 16: 2008	0,664	6,460	8,056	9,583	10,938	12,009	13,813	17,983	31,681	38,489	45,533	48,878
Protein (g/ 100 g) ISO 8968-1/IDF 20-1:2014	3,139	3,488	3,437	3,385	3,331	3,287	3,230	3,064	2,528	2,247	1,950	1,783
Lactose, Mono (g/100 g) ISO 5765-2/IDF 79:2002										3,147		
Dry Matter (g/100 g) ISO 6731/IDF 21:2010			16,773	18,133		20,300	21,944	25,697	38,265	44,454	50,612	53,726

** Increase uncertainty (U×2)

Version 1

Date: 01-08-2024

Bente Østergaard

CRM Approver, Bente Østergaard

The results apply only the received batch

The results may not be reproduced, except in their entirety, without written permission from the Laboratory.

Contact: calibration@ftn.eurofins.com

Information, if any:

Informations according to the certified reference material:

Description

Fresh Bovine Milk from Danish Dairies.
The Samples are added preservative.

The certified values are calculated by accredited reference methods.

Because of the short validy period, the reference value is done by one laboratory, but in several determinations and batch.

Bias are checked by comparing with FTIR values from Eurofins Milk Testing.

Homogeneity are checked on Reference values and FTIR values.

Stability has been calculated for a representative batch.

Instruction for handling and use:

The Samples are set at room temperature for 60 min, then heated in water bath to 38°C for 20 min.
Gently invert the Samples 20 times and measure immediately on FTIR equipment
The Samples can't be reused

Sample size:

Tube with 50mL

Period of validity:

14 days at 4°C ± 2°C from production date.

Storage information:

Store at 4°C ± 2°C until use.

The Reference Laboratory have some approved modification:

Fat: Drying fixed time instead of to constant weight
Protein: The destruction don't start with low temperature, but starts at 420°C
Dry Matter: Using dispoible cups without lids instead of glass bowls with lids.

The Samples are calculated with following expanded uncertainty U%

	U%
Fat	1,5%
Protein	1,5%
Lactose	2,0%
Dry Matter	1,0%