

Certified Reference Values



Milk Testing

Production date: 05-10-2026

Batch: 41 - 2026

	Unhomogenised past. milk	Unhomogenised past. cream										
Type	Buttermilk	6,5%	8%	9,5%	11%	12%	14%	18%	32%	38%	46%	50%
Fat (g/100 g) ≤ 8%+9,5%: ISO 1211/IDF 1:2010 ≥ 9%: ISO 2450/IDF 16: 2008	0,615	6,371	8,129	9,606	11,192	11,803	13,619	17,878	32,106	38,342	45,407	49,434
Protein (g/ 100 g) ISO 8968-1/IDF 20-1:2014	3,346	3,741	3,670	3,603	3,537	3,459	3,393	3,216	2,760	2,331	2,006	1,915
Lactose, Mono (g/100 g) ISO 5765-2/IDF 79:2002										3,014		
Dry Matter (g/100 g) ISO 6731/IDF 21:2010			17,095	18,386		20,295	21,961	25,855	38,861	44,390	50,930	54,711

** Increase uncertainty (U×2)

Version 1

Date: 08-10-2026

Sophie Markussen

CRM Approver, Sophie Skov Markussen

The results apply only the received batch

The results may not be reproduced, except in their entirety, without written permission from the Laboratory.

Contact: calibration@ftn.eurofins.com

Information, if any:

The shelf life of the 50% cream has been shortened to 12th October due to raw material supply issues

Informations according to the certified reference material:

Description

Bovine milk from Danish Dairies.
CRM has preservative added.

The certified values are calculated based on results from chemical analysis.

Because of the short validity period, reference analyses
is carried out at one laboratory only, but in several determinations
and batches for statistical robustness.

Homogeneity is controlled on data from both chemical and FTIR analysis.
Stability has been calculated for a representative batch.

*Following changes to the above methods have been made.

Fat: The drying is with a constant time instead of constant weight.

Protein: The destruction does not take place at lower temperatures, but starts at 420°C.

Dry matter: Single-use vials without lids are used instead of vials with lid, and the samples are not pre-dried before placed in the oven.

Instruction for handling and use:

Place the CRM at room temperature for 60 min, then heat in water bath to 38°C for 20 min.
Gently invert the CRM 20 times and measure immediately on FTIR equipment
Samples can't be reused. Once heated, please discard any remains after measurements have finished.

Sample size:

Approx. 50 mL.

Period of validity:

14 days at 4°C ± 2°C from production date.

Storage information:

Store at 4°C ± 2°C until use.

The Samples are calculated with following expanded uncertainty U%

	U%
Fat	2,0%
Protein	1,5%
Lactose	2,0%
Dry Matter	1,5%