

Certified Reference Values



Milk Testing

Production date: 5/5/2025

Batch: 19 - 2025

Type	Unhomogenised past. milk	Unhomogenised past. cream										
	Buttermilk	6.5%	8%	9.5%	11%	12%	14%	18%	32%	38%	46%	50%
Fat (g/100 g) <i>≤8%+9.5%: ISO 1211/IDF 1:2010</i> <i>≥9%: ISO 2450/IDF 16:2008</i>	0.683	6.536	8.157	9.295	11.138	12.322	13.996	17.803	31.992	37.295	44.770	50.167
Protein (g/100 g) <i>ISO 8968-1/IDF 20-1:2014</i>	3.480	3.689	3.617	3.562	3.486	3.415	3.335	3.168	2.536	2.298	1.937	1.846
Lactose, Mono (g/100 g) <i>ISO 5765-2/IDF 79:2002</i>										3.092		
Dry Matter (g/100 g) <i>ISO 6731/IDF 21:2010</i>			17.022	18.027		20.713	22.225	25.687	38.464	43.256	49.761	55.094

** Increase uncertainty (U×2)

Version 1

Date: 5/8/2025

Signature

CRM Approver, Signe Hannesbo Møller Nielsen

The results apply only the received batch

The results may not be reproduced, except in their entirety, without written permission from the Laboratory.

Contact: calibration@ftn.eurofins.com

Information, if any:

Informations according to the certified reference material:

Description

Bovine milk from Danish Dairies.

CRM has preservative added.

The certified values are calculated based on results from chemical analysis.

Because of the short validity period, reference analyses is carried out at one laboratory only, but in several determinations and batches for statistical robustness.

Homogeneity is controlled on data from both chemical and FTIR analysis. Stability has been calculated for a representative batch.

*Following changes to the above methods have been made.

Fat: The drying is with a constant time instead of constant weight.

Protein: The destruction does not take place at lower temperatures, but starts at 420°C.

Dry matter: Single-use vials without lids are used instead of vials with lid, and the samples are not pre-dried before placed in the oven.

Instruction for handling and use:

Place the CRM at room temperature for 60 min, then heat in water bath to 38°C for 20 min.

Gently invert the CRM 20 times and measure immediately on FTIR equipment

Samples can't be reused. Once heated, please discard any remains after measurements have finished.

Sample size:

Approx. 50 mL.

Period of validity:

14 days at 4°C ± 2°C from production date.

Storage information:

Store at 4°C ± 2°C until use.

The Samples are calculated with following expanded uncertainty U%

	U%
Fat	1.5%
Protein	1.5%
Lactose	2.0%
Dry Matter	1.0%