

**Proficiency test SPIL-4
(2020)**

Nitrogen parameters in wastewater (effluent)

Proficiency test SPIL-4 (2020) Quality Documentation

Eurofins Miljø A/S
Smedeskovvej 38
DK-8464 Galten
Denmark

Tlf: +45 7022 4266
e-mail: proficiency@eurofins.com
Web: www.eurofins.dk

January 2021

Client	Client's representative
Environmental laboratories	

Project	Project No
Proficiency test SPIL-4 (2020)	20405-04
Authors	Date
Rikke Mikkelsen	2021-01-26
	Approved by
	Peter Rerup
1	Quality Documentation Report
	FYE3 DPRE 2021-01-26
Revision	Description
	By Approved Date
Key words	Classification
Analytical quality, assigned value, precision, trueness, homogeneity, stability, total nitrogen, ammonium, nitrite+nitrate, conductivity, pH, wastewater	☒ Open ☐ Internal ☐ Proprietary

Distribution
Available online: www.eurofins.dk
Eurofins: Rikke Mikkelsen, Peter Rerup

CONTENTS

1	INTRODUCTION.....	1
2	FEATURES OF THE PROFICIENCY TEST	2
3	HOMOGENEITY AND STABILITY OF SAMPLES.....	4
4	CONCLUSION	5
5	REFERENCES	6
ANNEX A	LIST OF PARTICIPANTS.....	8
ANNEX B	SAMPLE PREPARATION	10
ANNEX C	CONTROL OF SPIKE VALUES	11
ANNEX D	CONTROL OF RECOVERY.....	17
ANNEX E	CONCENTRATION LEVEL.....	23
ANNEX F	HOMOGENEITY AND STABILITY	24

1 INTRODUCTION

A proficiency test on the analysis of nitrogen parameters in wastewater was conducted on 12 November. The proficiency test was organised by Eurofins Miljø A/S.

The present report contains Eurofins' documentation for the quality of the proficiency test. Results of the proficiency test including data from participating laboratories and statistical analysis of these data were issued in a report to all participants /1/ on 14 December 2020.

2 FEATURES OF THE PROFICIENCY TEST

Participants in the proficiency test were a total of 52 laboratories from Denmark, Norway and Sweden. A list of participants is shown in Appendix A.

The closing date for submission of results was 26 November. All participants except laboratory no. 6 and 25 had submitted their results before the dead-line.

2.1 Sample preparation

The parameters covered in the proficiency test are listed in Table 2 as are the abbreviations used in this report.

Four samples were dispatched for the proficiency test. The samples were sample pairs covering the parameters as described in Table 1. The matrix of the samples represented wastewater, in this case effluent. Sample preparation is described in Appendix B.

Table 1 Samples in the proficiency test

Sample name	Parameters
A1/B1	TN, NH ₄ , NO ₂₊₃ , NO ₃ and γ_{25}
A2/B2	pH

2.2 Statistical analysis of participants' data

A split-level design was used. The data analysis was performed in accordance with ISO 5725: "Accuracy (trueness and precision) of measurement methods and results" (2019) /2/ and as described in detail in Spliid (1992) /3/. A short introduction to the statistics and a list of symbols and abbreviations used is given in Eurofins document "Schedule for a proficiency test", which is available at Eurofins' home page /4/.

The statistical model used is based on the assumption that the variances for the two samples in a sample pair are identical. The assumption was tested (F-test, 95% confidence level) and the result was that the two variances may be assumed to be identical for all parameters

2.3 Assigned and spike value

An overview of the concentrations in the samples (the assigned values) and the difference in concentration between the two samples of a sample pair (spike value) are shown in Table 2 compared to the range of concentrations normally encountered in effluent. The table also gives the expanded uncertainty of the assigned values.

Table 2 Assigned and spike value

Parameter	Abbreviation	Unit	Typical Range	Assigned value	Uncertainty of assigned value	Spike value
Total nitrogen	TN	mg/L N	2-10	5.3	0.16	0.9
Ammonium	NH ₄	mg/L N	0.1-2	1.09	0.015	0.26
Nitrite+nitrate	NO ₂₊₃	mg/L N	1-5	2.99	0.035	0.58
Nitrate	NO ₃	mg/L N	1-5	3.07	0.029	0.58
Conductivity	γ_{25}	mS/m	50-300	68	1.0	1
pH	pH		6-9	7.79	0.040	0

2.3.1 Assigned and spike values

The content of each parameter in each sample is given an assigned value for the sample with the lower content and a spike value, the spike value being the difference in concentration between the two samples of the sample pair.

In order to ensure optimal use of the data, the assigned value is calculated as the average of the median for both samples in the sample pair after subtraction of the spike value. The spike values are calculated from sample preparation except for γ_{25} where the spike value is the difference between median values for the two samples in the sample pair.

The assigned value for TN is operationally defined and is a consensus value based upon the median for method no. 1, 2 and 3 (TN). A list of method identification numbers is found in the report to participants /1/. Assigned values for NH_4 , NO_{2+3} , NO_3 , conductivity and pH are consensus values for all laboratories based on the median.

2.3.2 Test of spike values

A comparison was made (t-test, 95% confidence level) between the spike value and the difference in concentration between the two samples in the sample pair found from the laboratories' results, see Appendix C. The test showed no significant difference between the two for most parameters.

The test revealed a significant difference between the two for γ_{25} . However, the difference is numerically small and has insignificant influence on the general quality of analyses estimated from the data as well as on the evaluation of accuracy of participating laboratories.

2.3.3 Test of assigned values

The assigned value and the average of the results obtained from all laboratories were also compared (t-test, 95% confidence level), see Appendix D. The test showed for most parameters no significant difference between the two and the control of assigned value at Eurofins confirmed the value (Appendix E). The test revealed a significant difference between the two for γ_{25} . Average recovery was 97.7 %. The difference could be attributed to influence from a few laboratories. The test was repeated after exclusion of the results from laboratory no. 11 and 27 and now showed no significant difference. The assigned value is therefore kept unchanged.

3 HOMOGENEITY AND STABILITY OF SAMPLES

The homogeneity and stability of samples was/were tested using the following parameters as indicators:

NH ₄	Combined homogeneity and stability test
pH	Combined homogeneity and stability test

The results of control measurements are shown in Appendix F. The appendix also gives the results of the statistical evaluation of the control data. The data are analysed by analysis of variance (ANOVA) giving:

1. the standard deviation/variance for replicates (the contribution from analytical variability),
2. the between bottle standard deviation/variance (the contribution from heterogeneity) and
3. the between days concentration difference (the contribution from instability).

Homogeneity is evaluated by comparing the between bottle variance to $0.3 \cdot \hat{\sigma}$ specified by the Danish EPA /5/, whereas the stability is evaluated by comparing the concentration change of the samples to $0.3 \cdot \hat{\sigma}$. This test ensures that heterogeneity and instability will not have negative influence on the evaluation of participant performance /6/.

The appendix also shows the standard deviation within and between laboratories from the proficiency test to allow comparison between tests performed and average quality from participating laboratories.

The tests for homogeneity show that the samples are homogeneous.

The tests for stability show that the samples are unstable. The result for NH₄ decreases from production week to the date of the proficiency test. The results found on the date for the proficiency test deviated less than $0.3 \cdot \hat{\sigma}$ from the nominal found using the consensus values for all laboratories based on the median. It is therefore concluded that the above found instability had no effect on the evaluation of the participants' results. The change in pH is insignificant compared to the standard deviation of the parameter. Since the proficiency test is set to a specific date the change in pH is estimated not to influence the evaluation of the participating laboratories.

4 CONCLUSION

The quality control performed, including test of sample homogeneity as well as test of recovery of spike and assigned values, shows that the samples and their assigned values are suitable for testing the proficiency of the participating laboratories for all parameters. The results are also suitable for estimation of the general quality of analyses among all participating laboratories.

For γ_{25} the participants could not recover the spike value or the assigned value. The difference between the calculated spike value and that found by the participants is small and the influence on evaluation of participant performance or estimation of general quality of analyses is insignificant. Also the scrutiny of the combined evidence gave the conclusion that the assigned value is correct. The assigned value is therefore kept unchanged and it is recommended as the basis for evaluation of participating laboratories.

The test performed at the laboratory making Proficiency testing at Eurofins found instability in the tested parameters. The change in parameters are larger than the acceptable limit $0.3 \cdot \hat{\sigma}$, but as described in the above chapter the change is estimated to have insignificant influence on the evaluation of the participants' performance.

5 REFERENCES

- /1/ Eurofins A/S, *Proficiency test SPIL-4 (2020)*, Report to participants, December 2020.
- /2/ ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results – Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*, 2019.
- /3/ Spliid, H., *Procedure and analysis of data for proficiency tests and environmental analyses*, Report to Danish Environmental Protection Agency, 1994 (in Danish).
- /4/ Eurofins A/S, *Schedule for a proficiency test*, document may be downloaded from www.eurofins.dk/proficiencytest.
- /5/ Ministry of Environment regulation no. 1770 on *quality criteria for environmental measurements*, 28 October 2020 (in Danish).
- /6/ ISO 13528, *Statistical methods for use in proficiency testing by interlaboratory comparison*, 2015.

A N N E X E S

ANNEX A LIST OF PARTICIPANTS

Laboratory	Town	Country
Spildevandslaboratoriet (2 participants)	Esbjerg	Denmark
AquaDjurs A/S	Grenaa	Denmark
Provas - Haderslev Forsyningservice A/S	Haderslev	Denmark
Hillerød Spildevand	Hillerød	Denmark
Holbæk RA, Fors Spildevand Holbæk	Holbæk	Denmark
Holstebro Renseanlæg, Vestforsyning Spildevand	Holstebro	Denmark
Kerteminde Forsyning - Spildevand A/S	Kerteminde	Denmark
BlueKolding A/S	Kolding	Denmark
Biofos A/S	København K	Denmark
Klarforsyning, Køge-Egnens Renseanlæg	Køge	Denmark
Melby Renseanlæg	Liseleje	Denmark
CP Kelco, QC Analyselab.	Ll. Skensved	Denmark
Nyborg Forsyning & Service	Nyborg	Denmark
NK-Forsyning A/S	Næstved	Denmark
VandCenter Syd A/S	Odense C	Denmark
Vandmiljø Randers A/S	Randers SØ	Denmark
Bjergmarken RA, Fors Spildevand Roskilde	Roskilde	Denmark
Rønne Renseanlæg	Rønne	Denmark
Ringkøbing-Skjern Forsyning A/S	Skjern	Denmark
RGS Nordic A/S	Skælskør	Denmark
Svendborg Spildevand A/S, Egsmade/Svenborg Centralrenseanlæg	Skårup Fyn	Denmark
SK Spildevand A/S, Slagelse Renseanlæg (5 participants)	Slagelse	Denmark
Eurofins Miljø A/S (2 participants)	Vejle	Denmark
Vejle Spildevand	Vejle	Denmark
Aalborg Forsyning, Renseanlæg Øst	Aalborg Øst	Denmark
SYNLAB AS (Hamar)	Hamar	Norway
VIVAB	Falkenberg	Sweden
Preem AB Göteborg	Göteborg	Sweden
St1 Refinery AB	Göteborg	Sweden
GRYAAB AB	Göteborg	Sweden
Hallsta Pappersbruk	Hallstavik	Sweden
AB Lennart Månsson International	Helsingborg	Sweden
Kalmar Vatten AB, VA-lab	Kalmar	Sweden
Kristianstad Kommun	Kristianstad	Sweden
Yara AB	Köping	Sweden
Käppalaförbundet	Lidingö	Sweden
Preemraff Lysekil	Lysekil	Sweden
Mjölby Kommun	Mjölby	Sweden
Motala Kommun, Tekniska Förvaltningen, Vatten och Avfall	Motala	Sweden
Fiskeby Board AB	Norrköping	Sweden

Laboratory	Town	Country
Holmen Paper AB, Bravikens Pappersbruk	Norrköping	Sweden
Nodra AB, Laboratoriet Slottshagens Reningsverk	Norrköping	Sweden
Ernemar Laboratorium	Oskarshamn	Sweden
Perstorp Specialty Chemicals AB	Perstorp	Sweden
Smurfit Kappa Piteå	Piteå	Sweden
Vallviks Bruk AB	Vallvik	Sweden
Västervik Miljö & Energi AB, Vattenlaboratoriet	Västervik	Sweden
Växjö Kommun, Tekniska Förvaltningen/VA-Lab.	Växjö	Sweden
Perstorp Oxo AB	Ödsmål	Sweden
Östersunds Provningslaboratorium, Gövikens reningsverk	Östersund	Sweden

ANNEX B SAMPLE PREPARATION

Stock solution	Prepared from	Concentration
Stock TN	3.000 g Disodium edetate, 2H ₂ O milli-Q water up to 1000.0 g	TN: 225.8 mg/kg N
Stock NH ₄	1.000 g Ammonium chloride (NH ₄ Cl) milli-Q water up to 1000.0 g	NH ₄ : 261.9 mg/kg N
Stock NO ₃	2.500 g Potassium nitrate (KNO ₃) milli-Q water up to 1000.0 g	NO ₃ : 346.4 mg/kg N

Sample	Sample prepared from	TN mg/L N	NH ₄ mg/L N	NO ₃ mg/L N	γ ₂₅ mS/m
A1	10.00 g stock TN 25.00 g stock NH ₄ 60.00 g stock NO ₃ Sample B1 up to 34.00 kg	0.997 · (a+1.39) + 0.0664	0.997 · (b+0.605) + 0.193	0.997 · (c+2.66) + 0.611	d
B1	400.0 g stock TN 150.0 g stock NH ₄ 500.0 g stock NO ₃ filtered water from Vejen sewage treatment plant up to 65.0 kg	a+1.39	b+0.605	c+2.66	e

Sample	Sample prepared from	pH
A2/B2	Filtered water from Vejen sewage treatment plant.	f

ANNEX C CONTROL OF SPIKE VALUES

Total Nitrogen, mg/L N

Control of differences within sample pairs

Laboratory	Difference AB
1	-0.110
2	-0.200
3	-0.000
4	-0.010
5	0.190
7	-0.020
8	0.050
9	-0.650
10	0.060
11	-0.100
12	-0.070
13	0.220
14	0.371
15	-0.000
16	0.540
17	-0.046
18	-0.050
19	0.180
20	-0.140
21	0.310
22	-0.390
23	0.010
24	0.000
26	-0.030
27	-0.018
28	-0.090
29	-0.090
30	0.260
31	0.100
32	0.020
33	-0.072
34	-0.210
35	0.070
36	-
37	0.165
38	-0.080
39	-0.700
40	-0.050
41	0.120
42	-0.010
43	-
44	0.040
45	-0.050
46	-0.060
47	0.170
48	-0.240
49	-0.090
50	-0.080
51	-0.150
52	0.040

No of labs., p	48
No of repl., n	2
d	-0.019
s ²	0.045
s	0.212
$t = \sqrt{p} \cdot (d/s)$	-0.6047
Sign. level, p(t)	0.5483

No test statistics were found to be significant

Ammonium, mg/L N

Control of differences within sample pairs

Laboratory	Difference AB
1	-0.010
2	-0.000
3	-0.010
4	0.010
5	0.040
7	-0.070
8	-
9	0.060
10	0.010
11	-
12	-0.010
13	0.000
14	0.019
15	-0.010
16	-0.020
17	-
18	-0.050
19	-0.010
20	0.000
21	0.000
22	0.018
23	-
24	0.050
26	-0.020
27	-0.015
28	-0.000
29	0.000
30	-0.020
31	-0.010
32	-0.010
33	0.010
34	0.000
35	-0.030
36	-
37	-
38	0.010
39	0.010
40	-0.020
41	-0.010
42	0.010
43	-0.060
44	0.040
45	0.000
46	0.010
47	0.078
48	0.000
49	-0.010
50	0.060
51	-0.030
52	-0.020

No of labs., p	44
No of repl., n	2
d	-0.000
s ²	0.001
s	0.029
$t = \sqrt{p} \cdot (d/s)$	-0.0540
Sign. level, p(t)	0.9572

No test statistics were found to be significant

Nitrite+Nitrate, mg/L N

Control of differences within sample pairs

Laboratory	Difference AB
1	0.020
2	-
3	-
4	-0.010
5	0.030
7	-
8	-
9	-
10	-
11	-
12	-0.260
13	-
14	-0.209
15	-
16	-0.080
17	0.041
18	-
19	-
20	-
21	-
22	-
23	-
24	0.030
26	-
27	-0.038
28	-
29	-
30	-0.010
31	-
32	-
33	-
34	-
35	0.030
36	-
37	-
38	-
39	-
40	-
41	-
42	-
43	-0.030
44	-
45	-
46	-
47	-
48	-
49	0.000
50	-
51	-
52	-0.040

No of labs., p	14
No of repl., n	2
d	-0.038
s ²	0.008
s	0.091
$t = \sqrt{p} \cdot (d/s)$	-1.5507
Sign. level, p(t)	0.1450

No test statistics were found to be significant

Nitrate, mg/L N

Control of differences within sample pairs

Laboratory	Difference AB
1	-0.040
2	0.020
3	0.040
4	-0.010
5	-
7	-
8	-
9	-
10	0.010
11	-
12	-0.270
13	-0.000
14	-0.210
15	-0.020
16	-
17	-
18	-0.110
19	-0.020
20	0.020
21	-
22	-0.060 UG
23	-
24	-
26	0.000
27	-0.024
28	-0.540 UC
29	-0.040
30	-
31	-0.040
32	0.030
33	-
34	0.020
35	-
36	-
37	-
38	-0.040
39	0.060
40	0.060
41	-
42	-0.240
43	-
44	-0.070
45	-
46	-0.020
47	0.166
48	-0.050
49	-
50	-0.040
51	0.030
52	-

No of labs., p	28
No of repl., n	2
d	-0.028
s ²	0.008
s	0.091
$t = \sqrt{p} \cdot (d/s)$	-1.6451
Sign. level, p(t)	0.1116

No test statistics were found to be significant

UC denotes a Cochran outlier

UG denotes a Grubbs outlier

Conductivity, mS/m
Control of differences within sample pairs

Laboratory	Difference AB
1	-
2	-
3	-
4	-
5	-0.50
7	-
8	-
9	-0.40
10	-
11	-0.45
12	-0.80
13	-
14	-
15	-0.50
16	-
17	0.33
18	-
19	-
20	-0.20
21	-
22	-
23	-0.30
24	-
26	3.80
27	-0.27
28	-
29	-0.20
30	-
31	-
32	-0.40
33	-
34	-
35	-1.10
36	-0.30
37	0.00
38	-
39	-0.40
40	-0.50
41	-
42	-0.60
43	-
44	0.10
45	-
46	-
47	-0.40
48	-
49	-0.20
50	-
51	-
52	-

UC

No of labs., p	20
No of repl., n	2
d	-0.35
s ²	0.09
s	0.31
$t = \sqrt{p} \cdot (d/s)$	-5.1783
Sign. level, p(t)	0.0001 ***

* denotes that there is a significant difference
(t-test, 5%-level)

** denotes that there is a significant difference
(t-test, 1%-level)

*** denotes that there is a significant difference
(t-test, 0.1%-level)

UC denotes a Cochran outlier

Difference for sample pair AB is significantly different from 0,
and data should be corrected with the difference (in spike value),
during execution of Cochran's test.

pH,
Control of differences within sample pairs

Laboratory	Difference AB
1	0.120
2	-0.180
3	-0.070
4	-0.020
5	-0.060
7	-
8	-0.010
9	0.010
10	-0.010
11	-0.020
12	0.000
13	-0.070
14	-0.050
15	-0.030
16	-0.040
17	0.030
18	-0.240
19	-
20	0.030
21	0.000
22	-0.080
23	-0.020
24	0.050
26	-0.010
27	-0.087
28	-0.430 UC
29	0.090
30	-0.030
31	-
32	-0.020
33	-
34	-0.060
35	-0.020
36	0.070
37	0.010
38	0.100
39	-0.008
40	-0.190
41	-0.050
42	0.000
43	-0.470 UC
44	0.010
45	-0.170
46	0.020
47	-0.020
48	0.020
49	0.240
50	0.070
51	0.020
52	-0.100

No of labs., p	44
No of repl., n	2
d	-0.018
s ²	0.007
s	0.083
$t = \sqrt{p} \cdot (d/s)$	-1.4007
Sign. level, p(t)	0.1685

No test statistics were found to be significant
UC denotes a Cochran outlier

ANNEX D CONTROL OF RECOVERY

Total nitrogen, mg/L N

Control of recovery, average of results

Laboratory	Sample pair AB
1	5.025
2	4.930
3	5.160
4	5.275
5	5.215
7	5.090
8	5.195
9	5.955
10	5.420
11	5.530
12	5.085
13	5.350
14	4.950
15	5.370
16	4.960
17	4.992
18	5.425
19	5.410
20	5.210
21	5.555
22	4.955
23	4.948
24	4.800
26	5.475
27	4.855
28	5.225
29	5.195
30	5.570
31	5.350
32	5.400
33	4.312
34	5.785
35	5.345
36	-
37	5.269
38	4.590
39	4.310
40	5.405
41	4.580
42	5.185
43	-
44	5.640
45	5.525
46	5.230
47	5.067
48	5.350
49	5.225
50	6.240
51	5.645
52	5.090

No of labs., p	48
No of repl., n	2
m	5.222
s ²	0.135
s	0.367
Assigned value, μ	5.3
Recovery, %	98.5
$t = \sqrt{p} \cdot (m - \mu) / s$	-1.4673
Sign. level, p(t)	0.1489

No test statistics were found to be significant

Ammonium, mg/L N

Control of recovery, average of results

Laboratory	Sample pair AB
1	1.025
2	1.120
3	1.125
4	1.315
5	1.100
7	1.045
8	-
9	0.950
10	1.115
11	-
12	1.075
13	1.100
14	1.071
15	1.125
16	1.130
17	-
18	1.175
19	1.095
20	1.110
21	1.110
22	0.861
23	-
24	1.195
26	1.010
27	1.038
28	1.120
29	1.080
30	0.980
31	1.095
32	1.095
33	1.065
34	1.110
35	1.075
36	-
37	-
38	1.105
39	1.045
40	1.110
41	1.065
42	1.115
43	1.160
44	1.330
45	1.060
46	1.085
47	1.085
48	1.040
49	0.915
50	1.090
51	1.105
52	1.030

No of labs., p	44
No of repl., n	2
m	1.088
s ²	0.007
s	0.081
Assigned value, μ	1.09
Recovery, %	99.8
$t = \sqrt{p} \cdot (m - \mu) / s$	-0.2035
Sign. level, p(t)	0.8397

No test statistics were found to be significant

Nitrite+Nitrate, mg/L N

Control of recovery, average of results

Laboratory	Sample pair AB
1	3.010
2	-
3	-
4	3.045
5	2.945
7	-
8	-
9	-
10	-
11	-
12	2.870
13	-
14	3.216
15	-
16	2.620
17	2.938
18	-
19	-
20	-
21	-
22	-
23	-
24	3.025
26	-
27	3.006
28	-
29	-
30	2.915
31	-
32	-
33	-
34	-
35	3.025
36	-
37	-
38	-
39	-
40	-
41	-
42	-
43	3.235
44	-
45	-
46	-
47	-
48	-
49	2.980
50	-
51	-
52	2.980

No of labs., p	14
No of repl., n	2
m	2.986
s ²	0.021
s	0.147
Assigned value, μ	2.99
Recovery, %	99.9
$t = \sqrt{p} \cdot (m - \mu) / s$	-0.0912
Sign. level, p(t)	0.9287

No test statistics were found to be significant

Nitrate, mg/L N

Control of recovery, average of results

Laboratory	Sample pair AB
1	2.940
2	3.030
3	3.080
4	3.045
5	-
7	-
8	-
9	-
10	3.095
11	-
12	2.855
13	3.180
14	3.209
15	3.040
16	-
17	-
18	3.155
19	3.050
20	2.910
21	-
22	2.580 UG
23	-
24	-
26	2.980
27	2.994
28	0.880 UC
29	3.060
30	-
31	3.060
32	3.055
33	-
34	3.060
35	-
36	-
37	-
38	3.180
39	3.040
40	3.210
41	-
42	3.280
43	-
44	3.285
45	-
46	3.110
47	3.005
48	3.115
49	-
50	3.120
51	3.125
52	-

No of labs., p	28
No of repl., n	2
m	3.081
s ²	0.010
s	0.102
Assigned value, μ	3.07
Recovery, %	100.4
$t = \sqrt{p} \cdot (m - \mu) / s$	0.5727
Sign. level, p(t)	0.5716

No test statistics were found to be significant
 UC denotes a Cochran outlier
 UG denotes a Grubbs outlier

Conductivity, mS/m

Control of recovery, average of results

Laboratory	Sample pair AB
1	-
2	-
3	-
4	-
5	67.55
7	-
8	-
9	67.50
10	-
11	61.38
12	67.70
13	-
14	-
15	66.65
16	-
17	65.84
18	-
19	-
20	68.40
21	-
22	-
23	67.85
24	-
26	71.80
27	61.36
28	-
29	67.90
30	-
31	-
32	66.50
33	-
34	-
35	64.35
36	69.45
37	63.40
38	-
39	69.00
40	65.25
41	-
42	62.50
43	-
44	71.15
45	-
46	-
47	69.30
48	-
49	66.34
50	-
51	-
52	-

UC

No of labs., p	20
No of repl., n	2
m	66.47
s ²	7.38
s	2.72
Assigned value, μ	68
Recovery, %	97.7
$t = \sqrt{p} \cdot (m - \mu) / s$	-2.5214
Sign. level, p(t)	0.0208 *

* denotes that there is a significant difference
(t-test, 5%-level)

** denotes that there is a significant difference
(t-test, 1%-level)

*** denotes that there is a significant difference
(t-test, 0.1%-level)

UC denotes a Cochran outlier

pH,
Control of recovery, average of results

Laboratory	Sample pair AB
1	7.950
2	7.260
3	7.755
4	7.750
5	7.770
7	-
8	7.955
9	7.845
10	7.905
11	7.860
12	7.870
13	7.615
14	7.845
15	7.755
16	7.630
17	7.855
18	7.490
19	-
20	7.845
21	7.800
22	7.790
23	7.930
24	7.825
26	7.745
27	7.665
28	7.225 UC
29	7.825
30	7.925
31	-
32	7.760
33	-
34	7.710
35	7.930
36	7.695
37	7.775
38	7.850
39	7.615
40	8.165
41	7.885
42	7.810
43	7.445 UC
44	8.155
45	7.815
46	8.050
47	7.790
48	7.630
49	7.990
50	7.725
51	7.630
52	7.580

No of labs., p	44
No of repl., n	2
m	7.796
s ²	0.027
s	0.163
Assigned value, μ	7.79
Recovery, %	100.1
$t = \sqrt{p} \cdot (m - \mu) / s$	0.2412
Sign. level, p(t)	0.8106

No test statistics were found to be significant
UC denotes a Cochran outlier

ANNEX E CONCENTRATION LEVEL

Parameter	Unit	Sample	Bottle no.	I	II	Bottle Average	Sample Average	Assigned value	Spike	
									Measured	Assigned
Total nitrogen	mg/L N	A1	5	6,32	6,35	6,34	6,39	6,177	0,783	0,904
			42	6,40	6,43	6,42				
			75	6,39	6,42	6,41				
		B1	12	5,58	5,52	5,55	5,60	5,273		
			40	5,52	5,52	5,52				
			52	5,75	5,72	5,74				
Ammonium	mg/L N	A1	5	1,35	1,36	1,36	1,36	1,35	0,262	0,260
			42	1,37	1,37	1,37				
			75	1,36	1,36	1,36				
		B1	12	1,11	1,11	1,11	1,10	1,09		
			40	1,09	1,09	1,09				
			52	1,10	1,10	1,10				
Nitrate	mg/L N	A1	5	3,81	3,80	3,81	3,80	3,65	0,60	0,58
			42	3,79	3,80	3,80				
			75	3,81	3,81	3,81				
		B1	12	3,23	3,22	3,23	3,21	3,07		
			40	3,20	3,20	3,20				
			52	3,20	3,20	3,20				
pH		A2	15	7,86	7,88	7,87	7,85	7,79	-0,04	0,00
			33	7,84	7,85	7,85				
			53	7,83	7,84	7,84				
		B2	20	7,80	7,81	7,81	7,81	7,79		
			37	7,82	7,82	7,82				
			61	7,79	7,80	7,80				

ANNEX F HOMOGENEITY AND STABILITY

PT:	SPIL-4
Parameter:	NH4
Unit:	mg/L N
Sigma:	0,1

Responsible for tests: HBP/S7MS

Homogeneity test

Date: 2020-10-27

Sample	x(a)	x(b)	average	sd	sd^2
A1-2	1,40	1,40	1,40	0,000	0,000
A1-9	1,42	1,42	1,42	0,000	0,000
A1-20	1,42	1,43	1,43	0,007	0,000
A1-30	1,41	1,41	1,41	0,000	0,000
A1-35	1,41	1,42	1,42	0,007	0,000
A1-39	1,40	1,40	1,40	0,000	0,000
A1-47	1,41	1,41	1,41	0,000	0,000
A1-52	1,43	1,43	1,43	0,000	0,000
A1-58	1,44	1,44	1,44	0,000	0,000
A1-64	1,43	1,43	1,43	0,000	0,000
A1-70	1,42	1,43	1,43	0,007	0,000
A1-78	1,41	1,41	1,41	0,000	0,000

For homogeneity

General average (x)	1,42
Sample average sd (s _x)	0,013
Within-sample sd (s _w):	0,004
Between-samples sd (ss):	0,012
S _L in the Proficiency Test:	0,080
S _R in the Proficiency Test:	0,082

Stability test

Date: 2020-11-12

Sample	x(a)	x(b)
A1-5	1,35	1,36
A1-75	1,36	1,36
A1-42	1,37	1,37

For stability

General average (y):	1,36
/x-y/ =	0,056

Conclusions

ss = 0,012 0.3*sigma= 0,023	
/x-y/ = 0,056	
Analytical quality	Is s _w < 0,15*sigma YES
Homogeneity:	Is ss < 0.3*sigma? YES
Stability:	/x-y/ < 0.3*sigma? NO

PT: SPIL-4
Parameter: pH
Unit:
Sigma: 0,072

Responsible for tests: HBP/S7MS

Homogeneity test Date: 2020-10-27

Sample	x(a)	x(b)	average	sd	sd^2
B2-2	7,78	7,81	7,8	0,021	0,000
B2-6	7,78	7,80	7,8	0,014	0,00
B2-16	7,77	7,79	7,8	0,014	0,000
B2-21	7,77	7,78	7,8	0,007	0,000
B2-28	7,76	7,77	7,8	0,007	0,000
B2-35	7,76	7,78	7,8	0,014	0,000
B2-42	7,77	7,78	7,8	0,007	0,000
B2-47	7,77	7,79	7,8	0,014	0,000
B2-56	7,77	7,78	7,8	0,007	0,000
B2-63	7,78	7,78	7,8	0,000	0,000
B2-71	7,77	7,77	7,8	0,000	0,000
B2-77	7,77	7,77	7,8	0,000	0,000

For homogeneity

General average (x)	7,777
Sample average sd (s _x)	0,009
Within-sample sd (s _w):	0,011
Between-samples sd (ss):	0,0038
S _L in the Proficiency Test:	0,157
S _R in the Proficiency Test:	0,168

Stability test Date: 2020-11-12

Sample	x(a)	x(b)
B2-20	7,80	7,81
B2-37	7,82	7,82
B2-61	7,79	7,80

For stability

General average (y):	7,807
/x-y/ =	0,030

Conclusions

ss = 0,004 0.3*sigma= 0,02	
/x-y/ = 0,0295833	
Analytical quality	Is s _w < 0,15*sigma NO
Homogeneity:	Is ss < 0.3*sigma? YES
Stability:	/x-y/ < 0.3*sigma? NO