

Certified Reference Values



Milk Testing

Production date: 10-08-2026

Batch: 33 - 2026

	Unhomogenised past. milk	Unhomogenised past. cream										
Type	Buttermilk	6,5%	8%	9,5%	11%	12%	14%	18%	32%	38%	46%	50%
Fat (g/100 g) ≤8%+9,5%: ISO 1211/IDF 1:2010 ≥9%: ISO 2450/IDF 16: 2008	0,692	6,555	8,080	9,528	11,000	12,016	14,014	17,862	31,848	38,388	45,833	49,753
Protein (g/ 100 g) ISO 8968-1/IDF 20-1:2014	3,334	3,629	3,570	3,515	3,454	3,395	3,314	3,162	2,599	2,327	2,021	1,852
Lactose, Mono (g/100 g) ISO 5765-2/IDF 79:2002										2,984		
Dry Matter (g/100 g) ISO 6731/IDF 21:2010			16,996	18,324		20,409	22,336	25,514	38,301	44,539	49,790	53,908

** Increase uncertainty (U×2)

Version 1

The results apply only the received batch

The results may not be reproduced, except in their entirety, without written permission from the Laboratory.

Date: 13-08-2026

Sophie Markussen

CRM Approver, Sophie Skov Markussen

Contact: calibration@ftn.eurofins.com

Information, if any:

The fat analysis was performed by Eurofins Food & Feed Testing Sweden (Lidköping) according to accreditation Swedac 1977

Informations according to the certified reference material:

Description

Bovine milk from Danish Dairies.
CRM has preservative added.

The certified values are calculated based on results from chemical analysis.

Because of the short validity period, reference analyses is carried out at one laboratory only, but in several determinations and batches for statistical robustness.

Homogeneity is controlled on data from both chemical and FTIR analysis.
Stability has been calculated for a representative batch.

*Following changes to the above methods have been made.

Fat: The drying is with a constant time instead of constant weight.

Protein: The destruction does not take place at lower temperatures, but starts at 420°C.

Dry matter: Single-use vials without lids are used instead of vials with lid, and the samples are not pre-dried before placed in the oven.

Instruction for handling and use:

Place the CRM at room temperature for 60 min, then heat in water bath to 38°C for 20 min.
Gently invert the CRM 20 times and measure immediately on FTIR equipment
Samples can't be reused. Once heated, please discard any remains after measurements have finished.

Sample size:

Approx. 50 mL.

Period of validity:

14 days at 4°C ± 2°C from production date.

Storage information:

Store at 4°C ± 2°C until use.

The Samples are calculated with following expanded uncertainty U%

	U%
Fat	2,0%
Protein	1,5%
Lactose	2,0%
Dry Matter	1,5%